

I. TABLE 1
1 INDIVIDUAL EQUIPMENT RECORD (IER) – (see also section VI)

		X-Ray Equipment	Processor & Hardcopy device	CR Reader	Digital detectors	Film Viewer	Reporting Monitor
a)	Unit - make, model and system ID	X	X	X	X		
b)	Generator – make, model and serial number	X					
c)	Product Licence number, date of the latest licence & reference where a copy of the licence is kept	X		X			
d)	Date of installation	X	X	X	X	X	X
e)	Operator's manual – (Indication that the operator's manual is available and	X	X	X	X		
f)	Results of acceptance tests	X	X	X	X		X
g)	Results of routine quality control tests	X	X	X	X	X	X
h)	Date(s) of tube replacement(s)	X					

i)	Details of repairs/maintenance and/or modification(s). The licence holder must ensure that all the applicable test(s) are performed that could be affected by the aforementioned	X	X	X	X	X	X
j)	Should any of the tests in table 2 indicate non-compliance or should any problems be detected (indicated), the licence holder must implement corrective maintenance (repairs), followed by re-testing	X	X	X	X	X	X
k)	Details of the IB and person(s) that performed the test(s)	X		X	X	X	X

1 The **X** in each cell for each category of equipment (column 3 to 11), indicates which information must be available in the IER.